

VERIFICATION OF CALIBRATION REPORT

*of DataMaster cdm Breath Test Instrument
State of Alaska*

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130301</u> ✓
Name <u>CLAY FARIS</u>	ID# <u>6170</u> ✓	Date <u>2-1-11</u> ✓
A Agency <u>UNIVERSITY POLICE DEPT, FBKS (FBK7)</u>		Phone # <u>(907) 474-7721</u>
Instrument Location <u>UPD DATAMASTER ROOM, UPD, FBKS, AK</u>		
B Alco S/N <u>X172931</u> ✓		Target Value <u>.082</u> ✓ High Pressure <u>825 psi</u>
Alco Test Values		
<u>.085</u> ✓ 1st Alco	<u>.085</u> ✓ 2nd Alco	
Signature <u>[Signature]</u>		
(OVER)		

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 2ND day of MAR, 2011

[Signature]
Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office

(Notary Seal Stamp)



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130301

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130301

FEBRUARY 01, 2011

OPERATOR'S NAME:

FARIS/CLAY

OPERATOR'S NUMBER: 6170

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI :

VOC

D.L. #: VOC

DEPT/AGENCY: FBK7

CASE/REPORT:

TEST TYPE: U

ALCO TARGET VALUE: .082

ALCO S/N: X172931

--- BREATH ANALYSIS ---

.082 ADJUSTED FOR 29.16 in
ALCO TARGET .079 01:41
BLANK TEST .000 01:42
INTERNAL STANDARD VERIFIED 01:42
ALCO TV 29.16 in .085 01:43
BLANK TEST .000 01:43
SUBJECT SAMPLE .000 01:44
BLANK TEST .000 01:45
ALCO TV 29.16 in .085 01:45
BLANK TEST .000 01:46

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130301

FEBRUARY 01, 2011

TIME 01:37

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/28/01
HEATERS
SAMPLE CHAMBER: 48c
BREATH TUBE: 41c
BAROMETER: 29.18 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEF
GHIJKLMNOPQRSTUVWXYZ[\]^_`*abcdefghijklmnop
qrstuvwxyz{|}~